


AS/NZS 4308:2008 - Certificate of Compliance

**MicroScreen Cup – Manufactured by Ameditech, Inc
Lot No: 43009-1**

The MicroScreen Cup was tested in accordance with Appendix B – “Verification of performance of on-site devices around the cut-off”, of the Australian and New Zealand Standard AS/NZS 4308:2008.

The MicroScreen Cup was shown to satisfy the specified performance criteria when results are interpreted as described in the package insert (Number: 42141-ImmII Revision 0, September 2007).

Results of the verification procedure used to assess compliance of the MicroScreen Cup have been documented and are available to any requesting authority and any accrediting agency.



Dr. John H. Vine
Laboratory Director

17 September 2009